



アナログ温度	アナログ湿度	デジタル温度	デジタル湿度	#1V0.1温度	#1V0.1湿度	#2V0.2温度	#2V0.2湿度
30.0	61.0	28.6	63.0	28.2	60.4	28.9	60.4
26.0	39.0	26.1	47.0	24.2	43.0	25.3	43.5
24.0	40.0	24.3	48.0	23.0	43.1	24.1	44.4
23.0	43.0	23.3	49.0	22.5	44.6	23.4	45.8
22.3	46.0	22.7	53.0	22.3	47.3	23.2	48.2
22.2	48.0	22.4	54.0	22.3	49.2	23.1	51.0
22.0	50.0	22.4	54.0	22.3	50.4	23.1	52.2
22.0	52.0	22.2	57.0	22.4	51.3	23.2	52.6
22.3	55.0	22.4	58.0	22.4	53.6	23.3	54.9
22.0	49.0	21.8	55.0	21.8	50.3	22.5	51.6
22.0	51.0	21.6	57.0	21.8	51.4	22.4	53.2
22.0	52.0	21.6	57.0	22.0	52.2	22.5	54.1
22.0	53.0	21.7	58.0	22.0	52.8	22.6	54.3
22.0	56.0	21.9	61.0	22.5	54.5	23.0	56.3
21.8	54.0	22.0	56.0	22.0	54.1	22.8	55.1
21.8	50.3	21.4	56.0	21.5	51.6	22.2	52.9
21.5	50.8	21.1	56.0	21.3	51.8	22.0	53.1
21.0	52.0	21.0	57.0	21.4	52.5	22.0	54.1
21.2	53.0	21.0	58.0	21.4	52.7	22.1	54.1
21.6	58.0	21.3	58.0	22.3	56.2	21.9	54.6
21.2	53.0	21.0	57.0	21.4	53.2	22.1	54.0
21.0	53.0	20.9	58.0	21.3	53.1	22.0	54.0
21.0	54.0	20.8	58.0	21.4	53.4	22.1	54.0
21.0	54.0	20.9	59.0	21.5	53.7	22.2	54.7
21.0	54.3	21.0	59.0	21.5	54.2	22.2	55.0
21.0	55.0	21.1	60.0	21.7	54.2	22.3	55.2
21.8	56.0	21.3	61.0	21.9	54.4	22.5	55.6
22.0	57.0	21.4	62.0	22.1	55.5	22.7	55.9
22.0	58.0	21.7	62.0	22.3	56.6	22.9	57.4
22.0	59.0	21.9	63.0	22.4	57.2	23.2	57.7
22.5	59.0	22.2	63.0	22.7	57.9	23.3	58.8
22.5	60.0	22.3	64.0	22.8	58.4	23.4	59.4
22.5	58.0	22.3	61.0	22.6	57.9	23.3	58.0
22.5	58.0	22.2	62.0	22.5	57.1	23.2	56.7
22.5	58.0	22.2	62.0	22.5	57.1	23.2	57.1
24.0	79.0	24.0	76.0	25.2	67.7	25.4	70.3
25.0	77.0	25.5	74.0	26.2	68.9	26.6	71.2
26.0	73.0	26.4	73.0	26.6	68.4	27.1	70.5
27.0	72.0	26.9	73.0	27.1	68.1	27.5	69.9
27.0	70.0	27.3	73.0	27.3	68.1	27.7	69.5
27.0	70.0	27.5	73.0	27.5	67.8	28.0	69.1
27.5	69.0	27.8	72.0	27.5	67.8	28.0	68.8
27.5	69.0	27.9	72.0	27.7	67.6	28.2	68.5
27.5	69.0	28.0	69.0	27.7	67.8	28.2	68.4
28.0	69.0	28.0	69.0	27.7	67.6	28.2	68.3
28.0	69.0	28.0	69.0	27.7	67.8	28.3	68.0
28.0	67.0	27.9	70.0	27.2	67.0	27.9	66.8
28.0	66.0	27.7	70.0	27.0	67.1	27.7	67.1
27.5	66.0	27.6	70.0	26.9	67.5	27.7	67.4
last							

温度



湿度



アナログ湿度:

- エンペックス気象計株式会社
- スーパーEX温度・湿度計
- バイマテリアル式スーパーEX湿度センサ
- ±2%RH(35~75%)、±4%RH(左記以外)

デジタル湿度:

- セイコーロック株式会社
- 電波ロックNQ501S
- 使用センサ不明
- ±8%(25°Cのとき)
- 5分間隔計測
- 閉空間 (30cm四方段ボール箱)
- 空気強制循環
- 氷で温度を降下

グラフ 8

